

Certificate of Analysis

Trastuzumab ELISA AA105

REF	IG-AA105	LOT	AA105-106		29.06.2026	IVD	CE
------------	----------	------------	-----------	---	------------	------------	-----------

Standard Curve			
10X	1:10 Diluted Study Concentration (ng/mL)	Measured Value OD 450/620nm	Range OD
Standard A	200	3.763	≥ 1.500
Standard B	60	2.012	-
Standard C	20	0.670	-
Standard D	6	0.246	-
Standard E	0	0.016	< 0.100

The test results are valid, as long as all criterias are met.

This product has been tested successfully by the Quality Control Department and was released for sale according to existing specifications.

Validation Date (dd,mm,yyyy): 29.06.2024

QC: PASS

