



Quality Control Certificate

Antibody to Tocilizumab ELISA BB108

REF IG-BB108	LOT BB108-202	11.02.2024	IVD (€
--------------	---------------	------------	--------

Standards/Positive Control/Cut-Off				
	Ready-to-use Concentration (ng/mL)	Measured Value (OD450nm)	Range OD	
Standard A	100	1,973	>0,500	
Standard B	30	0,711	>Standard C	
Standard C	10	0,228	>Standard D	
Standard D	3	0,106	>Standard E	
Standard E	0	0,048	<0,100	
Positive Control	48 ng/mL (± 20%)	0,993		
Cut-Off (CO)			StE <co<std< td=""></co<std<>	

The test results are valid, as long as all criterias are met.

This product has been tested successfully by the Quality Control Department and was released for sale according to existing specifications.

Validation Date (dd-mm-yy):11.02.2022

QC: PASS

www.aybaytech.com