

Quality Control Certificate

Rituximab ELISA (mAb-based) AB106

REF	IG-AB106	LOT	AB106-108		12.06.2027	IVD	CE
------------	----------	------------	-----------	---	------------	------------	-----------

Standard Curve			
10X	1:10 Diluted Study Concentration (ng/mL)	Measured Value OD 450/620nm	Range OD
Standard A	200	2.842	≥ 1.500
Standard B	60	1.195	-
Standard C	20	0.409	-
Standard D	6	0.157	-
Standard E	0	0.043	≤ 0.100

The test results are valid, as long as all criterias are met.

This product has been tested successfully by the Quality Control Department and was released for sale according to existing specifications.

Validation Date (dd,mm,yyyy): 12.06.2025

QC: PASS

