

Quality Control Certificate

Cetuximab ELISA (mAb-based) AB112

REF	IG-AB112	LOT	AB112-106		14.07.2027	IVD	CE
------------	----------	------------	-----------	---	------------	------------	-----------

Standard Curve			
10X	1:10 Diluted Study Concentration (ng/mL)	Measured Value OD 450/620nm	Range OD
Standard A	200	3.232	≥ 1.500
Standard B	60	1.314	-
Standard C	20	0.541	-
Standard D	6	0.166	-
Standard E	0	0.019	≤ 0.100

The test results are valid, as long as all criterias are met.

This product has been tested successfully by the Quality Control Department and was released for sale according to existing specifications.

Validation Date: 14.07.2025

QC: PASS

