

Certificate of Analysis (CoA)

Antibody to Infliximab ELISA BB101

REF IG-BB101	LOT BB101-211	 04.07.2027 (dd,mm,yyyy)	IVD 
---------------------	----------------------	---	--

Standards/Positive Control/Cut-Off			
	Ready-to-use Concentration (ng/mL)	Measured Value (OD450nm)	Range OD
Standard A	30	1.120	>0,500
Standard B	10	0.406	>Standard C
Standard C	3	0.166	>Standard D
Standard D	1	0.119	>Standard E
Standard E	0	0.058	<0,100
Calibrator		0.060	<0,100
Positive Control	15.66 ng/mL	0.615	± 20%

The test results are valid, as long as all criterias are met.

This product has been tested successfully by the Quality Control Department and was released for sale according to existing specifications.

Validation Date: 04.07.2025

QC: PASS

